



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 9, 2012 INITIALS: Dec OWNER: SPOKANE COUNTY SHERIFF'S OFFICE  
DISTRICT: ADDRESS:  
LOCATION: FM SHOP COUNTY:  
MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII  
MODEL: 101 ID. NO: 26904 CAPACITY: 20000 lbs. SECTIONS:  
MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",  
Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Jan 9, 2012  
Print name: DAVID CROMER

State of Washington

County of Thurston

Signed and sworn to (or affirmed before  
me on Jan. 9, 2012)

Patricia A. Geppert

(Signature of notary public)

Patricia A. Geppert

Printed Name of Notary

My appointment expires Sept. 14, 2015

